



World Class Accreditation

The American Association for Laboratory Accreditation

# Accredited Laboratory

A2LA has accredited

## VERITEQ INSTRUMENTS, INC.

*Richmond, CANADA*

for technical competence in the field of

### Calibration

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005 *General Requirements for the Competence of Testing and Calibration Laboratories*. This laboratory also meets the requirements of ANSI/NCSL Z540-1-1994 and any additional program requirements in the field of calibration. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (*refer to joint ISO-ILAC-IAF Communiqué dated 8 January 2009*).

Presented this 25<sup>th</sup> day of August 2009.



  
\_\_\_\_\_  
Peter Abney

President & CEO  
For the Accreditation Council  
Certificate Number 2069.01  
Valid to September 30, 2011

*For the calibrations to which this accreditation applies, please refer to the laboratory's Calibration Scope of Accreditation.*

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005  
& ANSI/NCSL Z540-1-1994

VERITEQ INSTRUMENTS, INC.  
Suite 100, 13775 Commerce Parkway  
Richmond, B.C., Canada V6V 2V4  
Wlad Nefedow Phone: 604 273 6850

CALIBRATION

Valid To: September 30, 2011

Certificate Number: 2069.01

In recognition of the successful completion of the A2LA evaluation process, accreditation is granted to this laboratory to perform the following calibrations<sup>1</sup>:

I. Thermodynamics

Parameter/Equipment	Range	Best Uncertainty <sup>2</sup> (±)	Comments
Relative Humidity Measuring Equipment	(10 to 95) % RH	0.6 % RH	Thunder Scientific 2500ST_LT
Temperature Measuring Devices	-100 °C to 110 °C	0.02 °C	Hart Scientific 2560 with Hart Scientific 5626 PRT

<sup>1</sup> This laboratory is conditionally available for commercial calibration service.

<sup>2</sup> “Best Uncertainty” is the smallest uncertainty of measurement that a laboratory can achieve within its scope of accreditation when performing more or less routine calibrations of nearly ideal measurement standards of nearly ideal measuring equipment. Best uncertainties represent expanded uncertainties expressed at approximately the 95 % level of confidence, usually using a coverage factor of  $k = 2$ . The best uncertainty of a specific calibration performed by the laboratory may be greater than the best uncertainty due to the behavior of the customer’s device and to influences from the circumstances of the specific calibration.

